

Automated, High-Throughput LC-MS/MS Workflow for the Analysis of 25-Hydroxyvitamin D_{2/3} and 3-*epi*-25-Hydroxyvitamin D₃

Lewis Couchman¹, Caje Moniz¹, Sarah Robinson², Justin Brooking²

¹Department of Clinical Biochemistry, King's College Hospital - NHS Foundation Trust, London, UK

²Thermo Fisher Scientific, Hemel Hempstead, UK

Key Words

Vitamin D, 25-hydroxyvitamin D_{2/3}, 3-*epi*-25-hydroxyvitamin D₃, TurboFlow method, Transcend, TSQ Vantage, Versette, LC-MS/MS

Goal

Develop an automated, high-throughput LC-MS/MS workflow for the analysis of 25-hydroxyvitamin D_{2/3} and 3-*epi*-25-hydroxyvitamin D₃ in human serum for research laboratories.

Introduction

Analysis of total serum 25-hydroxyvitamin D_{2/3} (25OHD, shown in Figure 1) is performed routinely in many research laboratories. Demand for this analysis continues to grow, and liquid chromatography with tandem mass spectrometry (LC-MS/MS) is increasingly used for this purpose.¹

Compared with other sample preparation techniques, Thermo Scientific™ TurboFlow™ technology has been shown to significantly improve the removal of matrix components prior to LC-MS/MS analysis of 25OHD.² However, there are additional pre-analytical steps that must be performed, the most important of which are the complete removal of the analytes from the endogenous vitamin D binding protein and the addition of isotopically-labeled internal standards for quantitation. When performed manually, these steps can increase the total analysis time by approximately two hours for a batch of 96 samples. This application note presents a workflow that uses an automated liquid handling system to reduce the time required to prepare a 96 well plate for analysis to less than 20 minutes.

Further, MS/MS alone is an achiral technique. This can be problematic for some isobaric 25OHD metabolites, notably 3-*epi*-25-hydroxyvitamin D₃ (shown in Figure 1 as 3-*epi*-25OHD₃). For accurate LC-MS/MS analysis of 25OHD₃, LC-MS/MS extended chromatographic analysis times are needed to resolve 3-*epi*-25OHD₃. In this application note, multiplexing technology is used to maximize throughput of the chromatographic method used to resolve interfering 3-*epi*-25OHD₃.

Experimental

Sample Preparation

Human serum samples from the international Vitamin D External Quality Assessment Scheme (DEQAS, samples 404 and 405) were used for the analysis.

All liquid handling was carried out using a Thermo Scientific™ Versette™ automated liquid handling system. An overview of the liquid handling procedure is shown in Figure 2. The Versette system was fitted with a 96 channel pipetting head and Thermo Scientific™ D.A.R.T.'S™ 300 µL extended-tip disposable pipette tips. Calibration standards and quality controls (both from the Chromsystems MassChrom® 25-hydroxyvitamin D_{2/3} kit) and samples (100 µL) were transferred from decapped 1 mL Thermo Scientific™ Nunc™ Cryobank storage vials to a 96 well filter plate. Internal standard solution (25 µL, ²H₆-25OHD₃) and precipitation reagent (200 µL), both from Chromsystems, were then added separately from the reagent reservoirs. The filter plates were covered and mixed on a plate shaker (600 rpm, 10 min). Supernatants were collected into a microtitre plate by centrifugation (200 g, 3 min). The plate was sealed with an adhesive plate seal and transferred to a Thermo Scientific™ Transcend™ TLX-2 system for analysis. All of the consumables utilized in the process are listed in Table 1.




Figure 1. Structures of 25-hydroxyvitamin D_{2/3}

Thermo
SCIENTIFIC




Figure 2. Liquid handling procedure




Figure 3. Cryobank vials



Figure 4. 8 Channel handheld screw cap capper/decapper



Figure 5. Versette liquid handling system and one rack of D.A.R.T.'S tips



Figure 6. Transcend system and TSQ Vantage mass spectrometer

Table 1. Parts and consumables required for specific steps in the liquid handling procedure

Step Number	Description	Part Number
1	1 mL Cryobank vials (96 tubes per rack)	374088
2	8 channel handheld screw cap capper/decapper with stand for Thermo Scientific Nunc-style tubes	4105NUN
3	Thermo Scientific™ Matrix™ 96 deep-well automation reservoir	1064-15-6
4	Filter plate	Chromsystems
2–8	Thermo Scientific Versette system - Base unit - 96 and 384 channel pipetting module - 6 position stage - 96 channel air disp. pipetting head, volume 5–300 μ L	650-01-BS 650-02-NTC 650-03-SPS 650-06-96300
4–7	Tips – 300 μ L extended-length D.A.R.T.'S	5536
8	Titer plate shaker	4625-1CEQ
9	Thermo Scientific™ Heraeus™ Labofuge™ 400 with microplate rotor package	75008371
10	Nunc pierceable 96 well cap mats	276002
	Nunc polypropylene microplate - 96 well (1.0 mL)	260252

Liquid Chromatography

Sample supernatants (100 μ L) were injected onto a TurboFlow XL C18 column (50 x 0.5 mm i.d.) under turbulent flow conditions (2 mL/min). Retained analytes were back-flushed from the TurboFlow column using elution solvent stored in a holding loop and focused onto a Thermo Scientific™ Accucore™ PFP analytical column (2.6 μ m particle size, 50 x 2.1 mm i.d.) maintained at 40 °C. During isocratic elution (0.40 mL/min) from the analytical column, the TurboFlow column was back-flushed with eluent C and the elution solvent loop refilled. Eluent flow was diverted to waste for 8 minutes following each injection onto the TurboFlow columns. The system was then re-equilibrated prior to the next injection.

System eluents (Fisher Chemical™ brand) were as follows:

Loading and Eluting Pumps A:	0.1 % (v/v) aqueous formic acid
Loading and Eluting Pumps B:	0.1 % (v/v) formic acid in methanol
Loading Pump C:	Acetone/2-propanol/acetonitrile (1:2:2 v/v/v)

Mass Spectrometry

Mass spectrometry was carried out in positive ionization mode using atmospheric pressure chemical ionization (APCI) on a Thermo Scientific™ TSQ Vantage™ triple-stage quadrupole LC-MS/MS system. Selected-reaction monitoring (SRM) transitions did not include water-loss fragmentations.³ MS/MS data were acquired for 5 minutes per analysis to allow multiplexing. Total LC time was 14 minutes, as shown in Figure 7, which means the TLX-2 system could do one analysis every 7 minutes).

Results and Discussion

As shown in Figure 7, retention times were 10.94, 11.47, and 11.82 minutes for 25OHD₃, 3-*epi*-25OHD₃, and 25OHD₂, respectively. Total analysis time was 14 minutes, including column re-equilibration. Sample 405 was correctly found to contain 3-*epi*-25OHD₃. This compound would have been misidentified as additional 25OHD₃ if a C18 analytical column, which does not resolve the epimer well, had been used as part of the LC-MS/MS method.




Figure 7. Chromatograms showing resolution of 3-*epi*-25OHD₃ from 25OHD₃ in DEQAS samples 404 and 405. 25OHD₂ chromatograms not shown. Sample 405 was prepared by addition of 3-*epi*-25OHD₃ to sample 404.⁴ Yellow highlighted area corresponds to data window for multiplexing.

Conclusion

The Versette automated liquid handling system reduced the time required to prepare 96 samples from two hours to less than 20 minutes – a dramatic reduction of 80%. In addition, the Versette system reduced the number of manual pipetting steps from as many as 864 to none. The workflow minimized manual errors, increased method precision, and reduced the risk of repetitive strain injury in research laboratories.

Using multiplexing on the Transcend TLX-2 and an analytical column using a pentafluorophenyl stationary phase, the method resolved 3-*epi*-25OHD₃, an interferent in most LC-MS/MS 25OHD₃ methods, without significantly decreasing chromatographic throughput.

The same basic workflow employing automated liquid handling and automated online sample preparation can be used by research laboratories for the analysis of other compounds including mass spectrometric immunoassay-selective (MSIA) assays to measure parathyroid hormone (PTH) and vitamin D-binding protein.

Acknowledgements

Thanks to Graham Carter (DEQAS) for permission to quote DEQAS data.

References

1. El-Kouhry, J.M.; Reineks, E.Z. and Wang, S. *Clin. Biochem.* **2011**, *44*, 66-76.
2. Leaver, N.; Chihoho, B.; Robinson, S.; *Quantitative Analysis of 25-OH Vitamin D₃/D₂ Comparing 1D Chromatography, 2D Chromatography and Automated Online Separation*; Application Note 526, AN63358_E 11/10S; Thermo Fisher Scientific; San Jose, CA, 2010.
3. Couchman, L.; Benton, C.M. and Moniz, C.F. *Clin. Chim. Acta* **2012**, *413*, 1239-1243.
4. Vitamin D External Quality Assessment Scheme (DEQAS) Distribution October 2011 Summary Report. London: Charing Cross Hospital; 2011.

For research use only. Not for use in diagnostic procedures.

www.thermoscientific.com

©2013 Thermo Fisher Scientific Inc. All rights reserved. MassChrom is a registered trademark of Chromsystems Instruments & Chemicals GmbH. ISO is a trademark of the International Standards Organization. All other trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries. This information is presented as an example of the capabilities of Thermo Fisher Scientific Inc. products. It is not intended to encourage use of these products in any manners that might infringe the intellectual property rights of others. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.



Thermo Fisher Scientific,
San Jose, CA USA is
ISO 9001:2008 Certified.

Africa-Other +27 11 570 1840	Europe-Other +43 1 333 50 34 0	Japan +81 45 453 9100	Spain +34 914 845 965
Australia +61 3 9757 4300	Finland/Norway/Sweden +46 8 556 468 00	Latin America +1 561 688 8700	Switzerland +41 61 716 77 00
Austria +43 1 333 50 34 0	France +33 1 60 92 48 00	Middle East +43 1 333 50 34 0	UK +44 1442 233555
Belgium +32 53 73 42 41	Germany +49 6103 408 1014	Netherlands +31 76 579 55 55	USA +1 800 532 4752
Canada +1 800 530 8447	India +91 22 6742 9434	New Zealand +64 9 980 6700	
China +86 10 8419 3588	Italy +39 02 950 591	Russia/CIS +43 1 333 50 34 0	
Denmark +45 70 23 62 60		South Africa +27 11 570 1840	

AN63703_E 02/13S

Thermo
SCIENTIFIC
Part of Thermo Fisher Scientific