

Motic®

**SMZ168 Series
Stereo Zoom Microscopes
Instruction Manual**

WWW.MOTIC.COM

MOTIC INCORPORATION LTD.



Parc d'Innovation - BP 50111 - F67403 Illkirch cedex

France

tél 03 88 67 14 14
fax 03 88 67 11 68
bioblock.vente@thermofisher.com
www.bioblock.com

Belgique / België

tél 056 260 260
fax 056 260 270
bioblock.belgium@thermofisher.com
www.bioblock.be


España

tfnº 91 515 92 34
fax 91 515 92 35
bioblock.ventas@thermofisher.com
www.es.fishersci.com

Table of Contents

Components	2
Introduction	3
Unpacking	4
Assembly	5
Operation	6
Maintenance	14
Troubleshooting	18
Moving the microscope	19
Repair	19
Warranty	19

Components



SMZ-168-BL

1. Eyepieces	9. Slide stage clips	17. Support collar
2. Diopter adjuster	10. Stage	18. Column
3. Eyepiece tubes	11. Base	19. On/Off switch
4. Prism housing	12. Power cord for incident illumination	20. Incident illumination switch (I)
5. Head holder locking screw	13. Head	21. Transmitted illumination switch (T)
6. Objectives	14. Zoom knob	22. Light intensity control
7. Incident illumination	15. Head holder	
8. Incident light adjustment screw	16. Focus knob	

Introduction

Thank you for your purchase of a Motic stereomicroscope. Motic stereomicroscopes are precision instruments, subjected to meticulous examination to reach you in perfect condition. Their design combines easy management and optimum functionality with minimum maintenance.

The information contained in this manual is likely to go beyond what the average user needs to know to use the stereomicroscope, however, it is provided to answer any queries that may arise.

Stereomicroscopes are used to study three-dimensional objects, examine small objects, or dissect biological specimens. They also permit the observation of slide specimens.

These instructions should be read carefully before operating the microscope. They will permit you to use your new stereomicroscope to its fullest capabilities. Terminology used to describe components and controls can be found in the diagram on page 2.

These instructions are based on the assembly and use of the SMZ-168 model with additional notes specifically for the other models in the series.

Unpacking

All components of the stereomicroscope have been carefully packed to make sure they reach you in perfect condition. We recommend that you do not discard any of the packing containers in case you need to return the microscope or store it for long periods of time; or should it become necessary to transport it to a technical service provider for any repair, or maintenance.

The box should contain the following components:

- **SMZ-168 (Binocular) :**

An illuminated base with a pole for a moveable head with illumination. A black and white opaque stage, a binocular head with eyepieces, eyepiece protectors, a frosted glass stage, a blue filter, a protective cover and a 1.5mm hexagonal key.

- **SMZ-168 (Trinocular) :**

An illuminated base with a pole for a moveable head with illumination. A black and white opaque stage, a trinocular head with eyepieces, eyepiece protectors, a frosted glass stage, a blue filter, a protective cover and a 1.5mm hexagonal key.

Remove and handle all components of the microscope with extreme care.

Avoid touching the lenses of the optical elements and keep them clear of dust, water or other contaminating agents as they could stain or damage the lens surface affecting the quality of the image.

Assembly

All the steps described for the assembly of the stereomicroscope must be undertaken with extreme care and without forcing the placement of the distinct parts and elements of the stereomicroscope.

- A. Place the base of the microscope (11) upright on a flat, stable and clean surface.
- B. Loosen the head holder locking screw (5) and place the head (13) in the head holder (15) with extreme care.
- C. Tighten the head holder locking screw (5).
- D. Connect the incident illumination to the power cord (12) located at the upper part of the column (18).

Warning: Before connecting the stereomicroscope to a power source, always check that the output voltage coincides with that of the stereomicroscope.

Operation

A. Starting Up

The stereomicroscope has two stages. One is a frosted glass stage and is used for the observation of microscope slides or samples that are thin or transparent, such as leaves, insect wings, etc. The black and white stage is used for non-transparent objects or for dissection. For the best contrast, choose the side of the stage best suited for your specimen.

Warning: Transmitted illumination can **ONLY** be used with the frosted glass stage. The heat generated by the transmitted illumination can melt or damage the black and white stage. Such damage would not be covered under warrantee.

- 1. Changing the stage.
 - a. Using your fingers, remove the stage (10).
 - b. To use the glass stage, insert the blue filter into the centre of the filter receptacle with the frosted surface facing down.
 - c. Replace the glass stage.
- 2. Before connecting the stereomicroscope to a power source, adjust the light intensity to its minimum (22). This must be repeated every time the stereomicroscope is turned on or off to prolong the life of the bulb.

Three light switches are located on the base of the microscope.

MAIN : The principal switch, which turns the whole unit on or off.

I : Turns incident light on or off (illumination from above).

T : Turns transmitted illumination on or off (illumination from below).

- 3. Turn the principal switch (19) to the ON position.

4. Turn on the incident illumination (20) or transmitted illumination (21) or both at the same time ("I" and "II") according to the sample being observed.
5. Light intensity should be adjusted according to the objective used or the type of sample being observed.
6. The angle of incident illumination can be adjusted by using the adjustment screw (8) which can be used to vary the orientation of the lens.

B. Interpupillary adjustment

1. Looking through the eyepieces (1), move the eyepiece tubes (3) by taking hold of the prism housing (4) and moving it outwards or inwards.
2. Interpupillary distance is correct when the two fields of view observed through both eyepieces appear as one complete field of view.
3. Interpupillary distance should be adjusted for each new user.

C. Focusing

1. Turn the zoom knob (14) to the lowest magnification, 0.75X.
2. Place a flat object or a microscope slide on the centre of the stage (10).
3. Turn the focusing knobs (16) to mid-focus range.
4. The head holder (15) is mounted on a column (18) on which it can be moved up or down depending on the size of the object to be focused on.
 - a. Support the head holder (15) with one hand without touching any lenses and with the other loosen the screw on the support collar (17). The head holder can then be slid to the base (11).
 - b. Without letting go of the head, loosen the head holder locking screw. (Fig.1)




Fig.1

- c. While looking through the eyepieces, move the head holder up or down until the object appears in focus.
 - d. Tighten the head holder locking screw. Do not let go of the head yet.
 - e. Slide the security collar up to the head holder and tighten the support collar screw. The head can then be released.
 - f. It is not necessary to adjust the head every time the sample is changed, only when unable to bring the image into focus.
5. Adjust the focus knobs (16) until the image appears sharp.

D. Diopter adjustment

Diopter adjustment collars are located on the eyepiece tubes. They are in normal position when the lower part of the collar is aligned with the sign marked on the eyepiece tube.

To compensate for differences in eyesight:

1. Using only the right eye, look through the right eyepiece (1) and adjust the focus.
2. Next, using the left eye, look through the eyepiece and adjust the focus by turning the diopter adjuster (2) located on the left-hand tube (3) until the image appears sharp. Do not use the focus knobs to adjust focus (16).

E. Changing magnification

1. Turn the zoom control (14) to the lowest magnification and bring the image into focus. Immediately switch to the highest magnification and do the same.
2. The stereomicroscope is now parfocal. However, some minor focus adjustments may have to be made when switching from a higher strength objective to a lower strength objective. Switching from a lower strength to higher strength objective should require no focus adjustment.

Attaching a photographic or video camera (Only for SMZ-168TL)

The SMZ-168 TL model comes equipped with a vertical image port on the head for the installation of a photographic reflex type camera or video camera, using the corresponding adapters.

At the back of the head there is a selection lever that transmits the image to the vertical port. When the lever is opened, the image is transmitted to the vertical port but cannot be seen through the right eyepiece.

- A. To attach a photographic camera, an adapter tube is needed. This tube includes a 2X lens which maintains parfocality between the binocular and vertical ports. This balance can be achieved providing that the parfocality has been adjusted correctly, as described in point "E. Changing magnification" in this manual.

The adapter tube has a T type thread at one end where adapter mounts for all reflex camera brand names on the market can be attached.

NB: The T type adapter mount is not included as it is an item specific to each brand of camera on the market

1. To connect the camera to the microscope, first remove the front lens of the camera and replace it with the corresponding T mount. Screw the adapter tube provided into the mount of the camera.
2. Loosen the knurled screw (Fig.2) situated on the side of the vertical port on the head sufficiently enough to remove the protective cover.




Fig.2

3. Insert the adapter tube with the camera already mounted into the vertical port.
If it does not insert easily, unscrew the knurled screw until the adapter tube fits perfectly.
 4. Re-tighten the knurled screw until the camera is secure.
 5. Operate the camera according to the manufacturer's instructions.
- B. To attach a video camera, an adapter tube is needed. This adapter tube includes a 0.5X lens which maintains parfocality between the binocular and vertical ports. This is providing that parfocality has been correctly adjusted according to the steps followed in point "E. Changing magnification" in this manual.

The adapter tube is equipped with a "C" thread and a "CS" ring which can adapt different kinds of video cameras.

To connect the camera to the microscope, screw the adapter tube into the video camera.

1. Loosen the knurled screw (Fig.2) on the side of the vertical port of the head sufficiently enough to remove the protective cover.
2. Insert the adapter tube with the camera already mounted into the vertical port. If it does not insert easily, unscrew the knurled screw until the adapter tube fits perfectly.
3. Re-tighten the knurled screw until the camera is secure.
4. Operate the camera according to the manufacturer's instructions.

If the image on the monitor appears out of focus when the objective is changed, it is possible that the CS mount is responsible. Replace, or readjust it to obtain parfocality.

Attaching the Motic K2401 Fluorescent Ring Illuminator

The SMZ-168 comes equipped with a screw type objective system at the bottom of the optical tube for the installation of auxiliary objectives, or the K2401 Fluorescent Ring Illuminator with corresponding adapter.

A. To attach an auxiliary objective, the auxiliary objective must be a screw mount type to ensure a secure mount.

1. To mount the auxiliary objective to the microscope, first remove the protective glass from the optical tube by unscrewing it in a counter-clockwise direction.
2. Screw on the appropriate auxiliary objective by turning the objective [threads facing up] clockwise to secure the objective in the optical tube. There is no need to replace the protective glass.
3. Adjust the working distance according to the magnification and properties of the objective in use by turning the coarse focus either clockwise or counter-clockwise until the sample is in focus. Refer to figure 3 for the corresponding working distances of specific objectives.

If the image appears out of focus or slanted when viewed after installing an auxiliary objective, it is possible the objective may be mounted incorrectly. Remove the objective and refasten with care to ensure proper alignment.



Fig.3: Auxiliary Objectives

Standard Objective	Auxiliary Objectives				
	0.3X	0.5X	0.75X	1.5X	2X
113mm	343mm	200mm	130mm	54mm	34.5mm

Fig. 3: Working Distances of Auxiliary Objectives

B. To attach the K2401 Fluorescent Ring Illuminator, an SMZ-168 ring adapter is needed. This ring adapter will replace the previously installed fastening mechanism.

NB: *The SMZ-168 ring illuminator adapter is not included in the K2401 Fluorescent Ring Illuminator kit as it is a specific model adapter.*

1. To connect the K2401 Fluorescent Ring Illuminator, first remove the previously installed mounting mechanism by turning the screw mount ring in a counter-clockwise direction while holding the locking ring located at the bottom of the light.
2. Remove the protective glass of the optical tube by unscrewing it in a counter-clockwise direction.
3. Place the ring illuminator underneath the optical tube. The ring illuminator is not secured yet. Secure the ring illuminator to the optical tube by screwing the SMZ-168 adapter ring, from the bottom of the ring illuminator, into the optical tube in a clockwise direction. The ring illuminator is secured when the adapter becomes tight.
4. Adjust the position of the ring illuminator to avoid interference with the focusing and zoom knobs by turning slightly with your hand.
5. Plug the ring illuminator into the nearest electrical outlet free from debris.
6. Turn the ring illuminator on by pressing the principal to "O" for on.
7. Adjust the microscope focusing knobs to achieve the desired illumination effect and clarity.

If the ring illuminator does not turn on, check whether the outlet has been switched on. If still no illumination, unplug the ring illuminator. Next, unscrew the SMZ-168 adapter ring by turning it counter-clockwise while safely holding the ring illuminator by the handle end. Once removed, inspect the bulb of the ring illuminator for signs of malfunction or burnout.

Maintenance

Warning: For your own safety switch off and disconnect the microscope from any electrical sources before attempting any maintenance to avoid risk of electrocution. Consult your distributor if any repair or maintenance is required for your microscope that does not appear in this instruction manual.

A. Optical maintenance

Do not attempt to disassemble any optical components. For any repair work not specified in this manual, consult the technical service provider responsible for your area.

Before cleaning the lens surface, remove any dust with a brush specifically used for cleaning lenses, or with low pressure compressed air found in any photography shop.

1. Cleaning the eyepieces
 - a. Do not remove the eyepieces (1) from the eyepiece tubes (3).
 - b. Clean the external surface by dampening the lens with one's breath.
 - c. Afterwards, dry the lens with special lens paper. Dry in circular movements from the centre of the lens outwards. Do not wipe the lens when already dry as it will scratch easily.
2. Cleaning the objectives
 - a. Do not remove objectives from the microscope.
 - b. Clean only the surface area. Use a soft cotton cloth dampened slightly with Xylene. Dry the lens afterwards with the same cloth.

B. Electrical maintenance

1. Changing the bulb

- *Changing the transmitted illumination bulb*

- Rest the stereomicroscope on its side being extremely careful, especially with the eyepiece (1) and the stage (10).
- Unscrew the 6 screws and open the cap on the base.
- With a cloth, carefully pull the bulb out by disconnecting it from the socket.
- Do not touch the new bulb with your hands. Use a clean cloth to insert the pins of the bulb into the socket.
- If the bulb is accidentally touched with bare hands it must be cleaned as this could affect the transmission of light and life span of the bulb.
- Close the cover on the base and screw down firmly.

- *Changing the incident illumination bulb*

- Unscrew the illuminator protection tube (7) by turning it in a counter-clockwise direction and carefully remove it.




Fig.4 Incident Light

- With a cloth, carefully pull the bulb out to disconnect it from the socket.

- If the bulb is accidentally touched with bare hands, it must be cleaned as this could affect the transmission of light and life span of the bulb.

- Carefully insert a new bulb and replace the protection tube.

2. Changing the fuse

- With a flat screwdriver, lightly press on the slot of the fuse holder cover and turn 1/4 in the direction marked by the arrow.
- Release pressure and completely remove the loosened cover.
- Remove the fuse from the removed cover and insert the new one, ensure that it is 0.5 Amps.
- Insert the cover.
- Repeat step (a.) but turning 1/4 in the opposite direction of the arrow.
The cover must be firmly closed.




Fig.5: Fuse

C. Mechanical maintenance

1. Adjusting the focus tension

Tension comes pre-adjusted by the factory. The best point of tension is that which permits the focus knobs to move as loosely as possible without the head sliding down from its own weight.

The tension adjustment collar for focusing is situated between the focusing knob (16) and the head holder (15).

- a. Loosen the screw located in the collar hole with the 2mm hexagonal key.
- b. To increase the tension, turn the collar counter-clockwise; to decrease it, turn the collar clockwise.
- c. Re-tighten using the hexagonal key.

Troubleshooting

Electrical Problems

Problem	Cause	Solution
The bulb does not work	• Outlet inoperative • Cable not connected • Bulb burnt out • Fuse blown • Wrong Bulb	• Have it repaired by a qualified specialised technician • Connect the cable to the power source • Replace bulb • Replace fuse • Replace with appropriate bulb
Bulb has short life span	Very high voltage	Reduce light intensity to the minimum before turning on or off the stereomicroscope
Bulb burns out immediately	Wrong bulb	Replace with the appropriate bulb
Bulb flickers	• The bulb is not inserted correctly into socket • Bulb on the point of burning out • Fuse cover closed incorrectly • Bad connection with power source	• Insert bulb correctly • Replace bulb • Close correctly • Have it repaired by a qualified specialised technician
Fuse blown quickly	Wrong fuse	Replace with appropriate fuse
Fuse blown instantly	Short circuit	Have it repaired by a qualified specialised technician

Image Quality

Problem	Cause	Solution
Poor resolution	• Eyepieces dirty • Objectives dirty	• Clean eyepieces • Clean objectives
Spots, or stains in field of view	Eyepieces dirty	Clean eyepieces

* NB. Stains in the field of view may also be attributed to dirt on the inside of the eyepiece. It is recommended therefore that the lens be cleaned by a recognised service technician.

Mechanical Problems

Problem	Cause	Solution
Microscope does not stay in focus	The head drops down	Adjust the tension of the coarse focus knob

Moving the microscope

- If possible, avoid moving the stereomicroscope.
- Carry the stereomicroscope in both hands. One hand should hold the stereomicroscope column (18), and the other should support it under the base (11).
- Maintain the stereomicroscope in a vertical position.

Repair

If the stereomicroscope needs repairing by authorised personnel, we recommend that it be stored in its polystyrene box and returned to the distributor. Attach a note with a description of the problem.

Warranty

All Motic microscopes are warranted against any manufacturing defect for a 5 year period. Damage occurring as the result of any unauthorised repair work, or occurring through misuse or modification of the microscope will not be included under the conditions of the warranty. Bulbs and fuses are not under warranty.

The warranty service is provided by Motic or its authorised distributors. Defective products will be repaired free of charge when returned to Motic or one of its distributors. Transport costs will be covered by the purchaser.

Owing to possible modifications and improvements in their manufacture, changes may occur to stereomicroscopes without prior notice.

Microscope body		SMZ-168	SMZ-168-60	SMZ-168 T
	Magnification	0.75x - 5x		
	Zoom ratio	6.7 : 1		
	Working distance	113mm		
	Tube inclination angle	35°	60°	35°
	Interpupillary distance adjustment	Diopter Adjustment both eyes: ±5 Interpupillary adjustment: 52mm to 79mm		
	Video camera adaptability	-	-	C-mount [CCD 0.3X / CCD 0.65X not included]
	Zoom Adjustment Knob	Left/right - single shaft horizontal knob Interpupillary distance high/low magnification stopper incorporated		
Auxiliary objectives		Mounted by screwing into thread at bottom of optical body		
Eyepiece		WF 10X/23 High Eyepoint		
Stand		168P	168L	
		Streamlined stand	Transmitted illumination stand	
	Frame Installation	Mounting diameter: 76mm		
	Focusing Adjustment	50mm vertical movement		
	Stage plate	Black & white Frosted glass plate	Black & white Frosted glass plate	
	Light source	Cold light illumination [optional] / Fluorescent ring illuminator attachable [optional]	Transmitted illumination: Halogen Reflected illumination: Halogen Input rating: 12V/10W	